



Product specification

Feline Panleukopenia Virus Antigen

| | |
|-----------------------------|--|
| Date: | 01/01/2022 |
| Code number: | Feline Panleukopenia Virus (FPV) antigen positive |
| Lot number: | F030122 |
| Expiry date: | 01/01/2023 |
| Physical state: | Freeze dried |
| Size: | 0,5ml Feline Panleukopenia Virus antigen |
| Titer: | 1:180 in ELISA |
| Titer determination: | ELISA according to EVL testkit no: F1003-AG01 |
| Preservative: | 0,01 % (Thiomersal) preservative added |
| Storage: | Divide into aliquots, after opening and reconstitution freeze at -20°C. Avoid repeated freezing and thawing |
| Applications: | Usable as Feline Panleukopenia Virus antigen for standardization procedures on different immunological platforms |
| Source: | Concentrated purified tissue culture supernatant |
| Purification: | Inactivated, concentrated, centrifugated, 0,2µ filtered, cross filtered supernatant |
| Production method: | Cells persistently infected with FPV clone |

The product listed herein is for research use only and is not intended for use in human or clinical diagnosis. Nothing disclosed herein is to be construed as a recommendation to use this product in violation of any patents. The information present above is believed to be accurate. However, said information and products are offered without warranty of guarantee since the ultimate conditions of use and the variability of the materials treated are beyond our control. We cannot be responsible for patent infringements or violations that may occur with the use of this product. No claims beyond replacement of unacceptable material or refund of purchase price shall be allowed.